

EC DECLARATION OF CONFORMITY

BOWA-electronic GmbH & Co. KG
Heinrich-Hertz Strasse 4-10
72810 Gomaringen / Germany

We hereby declare in sole responsibility that the medical devices listed on the following pages
are conform with the essential requirements of the following standard:

Medical Device Directive (Council Directive 93/42/EEC of 14 June 1993)

with applied standards:

EN ISO 14971 / EN 60601-1 / EN 60601-1-2 / EN 60601-2-2 / EN ISO 13485 / DIN EN 1041

CE Marking

Class IIb according to annex II CE 0123


Notified Body

TÜV Product Service
Ridlerstr. 65, D-80339 München
Nr. 0123

The declaration of conformity is valid until: 08.03.2022

Gomaringen, 08.03.2017

Quality Management /
Regulatory Affairs



(Wolf-Rüdiger Fritz)

REF	Product description	Class	Rule	UMDNS	GMDN
900-100	Electrosurgical unit ARC 100	IIb	9	11-490	11490
900-303	Electrosurgical unit ARC 303	IIb	9	11-490	11490
900-350	Electrosurgical unit ARC 350 (REF 900-350)	IIb	9	11-490	11490
900-351	Electrosurgical unit ARC 350 (REF 900-351)	IIb	9	11-490	11490
900-400	Electrosurgical unit ARC 400	IIb	9	11-490	11490